

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060319\
 Data File : VX010030.D
 Acq On : 03 Jun 2019 12:54
 Operator : JC/SP
 Sample : K3129-04DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW003-190528DL

Quant Time: Jun 04 04:11:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178311	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	285658	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	260466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	115620	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119017	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
35) Dibromofluoromethane	5.50	113	87796	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
50) Toluene-d8	8.71	98	362101	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
62) 4-Bromofluorobenzene	11.13	95	124002	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
Target Compounds						
4) Vinyl Chloride	1.41	62	2844	1.077	ug/l #	85
27) cis-1,2-Dichloroethene	4.60	96	130903	56.439	ug/l	87
65) Chlorobenzene	10.14	112	2297	0.421	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060319\
Data File : VX010030.D
Acq On : 03 Jun 2019 12:54
Operator : JC/SP
Sample : K3129-04DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW003-190528DL

Quant Time: Jun 04 04:11:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

