

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060319\
 Data File : VX010032.D
 Acq On : 03 Jun 2019 13:40
 Operator : JC/SP
 Sample : K3129-08DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012RW162-190529DL

Quant Time: Jun 04 04:15:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	173749	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	275194	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	254027	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113563	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113855	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
35) Dibromofluoromethane	5.49	113	84501	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
50) Toluene-d8	8.71	98	351378	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	11.13	95	119472	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
Target Compounds						
4) Vinyl Chloride	1.41	62	30783	11.963	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	79799	35.309	ug/l	88
44) Trichloroethene	7.21	130	1038	0.516	ug/l	82
65) Chlorobenzene	10.14	112	15450	2.905	ug/l	99
91) 1,2-Dichlorobenzene	12.39	146	1151	0.307	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060319\
Data File : VX010032.D
Acq On : 03 Jun 2019 13:40
Operator : JC/SP
Sample : K3129-08DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012RW162-190529DL

Quant Time: Jun 04 04:15:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

