

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060319\
 Data File : VX010034.D
 Acq On : 03 Jun 2019 14:27
 Operator : JC/SP
 Sample : K3131-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0528-190530

Quant Time: Jun 04 04:19:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	174695	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	278095	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	246042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	106806	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114506	46.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.80%	
35) Dibromofluoromethane	5.50	113	84444	48.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.60%	
50) Toluene-d8	8.71	98	347942	49.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.02%	
62) 4-Bromofluorobenzene	11.13	95	116128	45.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.08%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060319\
Data File : VX010034.D
Acq On : 03 Jun 2019 14:27
Operator : JC/SP
Sample : K3131-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0528-190530

Quant Time: Jun 04 04:19:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

