

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060319\
 Data File : VX010038.D
 Acq On : 03 Jun 2019 16:01
 Operator : JC/SP
 Sample : K2641-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-053119-B

Quant Time: Jun 04 04:46:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	284886	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	261063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118567	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
35) Dibromofluoromethane	5.50	113	88576	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
50) Toluene-d8	8.71	98	366248	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.13	95	122988	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
Target Compounds						
16) Acetone	2.44	43	9821	6.702	ug/l	100
20) Methylene Chloride	2.86	84	265918	118.492	ug/l	97
43) Isopropyl Acetate	6.45	43	11840	1.822	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060319\
Data File : VX010038.D
Acq On : 03 Jun 2019 16:01
Operator : JC/SP
Sample : K2641-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-053119-B

Quant Time: Jun 04 04:46:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

