

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060319\
 Data File : VX010039.D
 Acq On : 03 Jun 2019 16:24
 Operator : JC/SP
 Sample : K2641-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-053119-C

Quant Time: Jun 04 04:50:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180912	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	289579	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	264699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	117281	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119290	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
35) Dibromofluoromethane	5.50	113	88813	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	8.71	98	367725	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.13	95	124708	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
Target Compounds						
16) Acetone	2.44	43	18361	12.435	ug/l	99
20) Methylene Chloride	2.85	84	427788	189.193	ug/l	98
43) Isopropyl Acetate	6.45	43	12115	1.834	ug/l	100
95) Naphthalene	13.83	128	30046	3.490	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060319\
Data File : VX010039.D
Acq On : 03 Jun 2019 16:24
Operator : JC/SP
Sample : K2641-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-053119-C

Quant Time: Jun 04 04:50:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

