

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060319\
 Data File : VX010042.D
 Acq On : 03 Jun 2019 17:34
 Operator : JC/SP
 Sample : K3139-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1A-S

Quant Time: Jun 04 05:01:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	172946	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	276009	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	254213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113488	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114581	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
35) Dibromofluoromethane	5.50	113	84774	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	8.71	98	355505	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.13	95	123055	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
Target Compounds						
16) Acetone	2.44	43	10220	7.240	ug/l	99
20) Methylene Chloride	2.86	84	36386	16.833	ug/l	99
43) Isopropyl Acetate	6.45	43	11422	1.814	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060319\
Data File : VX010042.D
Acq On : 03 Jun 2019 17:34
Operator : JC/SP
Sample : K3139-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1A-S

Quant Time: Jun 04 05:01:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

