

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX060319\
 Data File : VX010025.D
 Acq On : 03 Jun 2019 10:28
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 03:59:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.611	0.678	-11.0	96	0.00
3 P	Chloromethane	0.821	0.704	14.3	85	0.00
4 C	Vinyl Chloride	0.740	0.671	9.3#	87	0.00
5 T	Bromomethane	0.449	0.378	15.8	88	0.00
6 T	Chloroethane	0.456	0.402	11.8	90	0.00
7 T	Trichlorofluoromethane	0.818	0.798	2.4	93	0.00
8 T	Diethyl Ether	0.385	0.338	12.2	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.496	-0.4	98	0.00
10 T	Methyl Iodide	0.551	0.494	10.3	79	0.00
11 T	Tert butyl alcohol	0.174	0.153	12.1	90	0.00
12 CM	1,1-Dichloroethene	0.506	0.489	3.4#	95	0.00
13 T	Acrolein	0.110	0.118	-7.3	96	0.00
14 T	Allyl chloride	1.231	1.195	2.9	95	0.00
15 T	Acrylonitrile	0.419	0.378	9.8	92	0.00
16 T	Acetone	0.408	0.406	0.5	100	0.00
17 T	Carbon Disulfide	1.653	1.458	11.8	93	0.00
18 T	Methyl Acetate	0.950	0.884	6.9	99	0.00
19 T	Methyl tert-butyl Ether	1.912	1.866	2.4	98	0.00
20 T	Methylene Chloride	0.625	0.584	6.6	96	0.00
21 T	trans-1,2-Dichloroethene	0.563	0.532	5.5	98	0.00
22 T	Diisopropyl ether	2.342	2.255	3.7	95	0.00
23 T	Vinyl Acetate	2.104	2.019	4.0	94	0.00
24 P	1,1-Dichloroethane	1.135	1.093	3.7	96	0.00
25 T	2-Butanone	0.636	0.596	6.3	94	0.00
26 T	2,2-Dichloropropane	0.858	0.872	-1.6	100	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.612	5.8	98	0.00
28 T	Bromochloromethane	0.589	0.550	6.6	98	-0.01
29 T	Tetrahydrofuran	0.408	0.369	9.6	91	0.00
30 C	Chloroform	1.023	0.998	2.4#	98	0.00
31 T	Cyclohexane	1.075	1.085	-0.9	94	0.00
32 T	1,1,1-Trichloroethane	0.833	0.850	-2.0	98	0.00
33 S	1,2-Dichloroethane-d4	0.706	0.646	8.5	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.314	0.315	-0.3	96	0.00
36 T	1,1-Dichloropropene	0.507	0.522	-3.0	99	0.00
37 T	Ethyl Acetate	0.707	0.700	1.0	91	0.00
38 T	Carbon Tetrachloride	0.438	0.477	-8.9	99	0.00
39 T	Methylcyclohexane	0.602	0.672	-11.6	100	0.00
40 TM	Benzene	1.505	1.562	-3.8	97	0.00
41 T	Methacrylonitrile	0.413	0.401	2.9	93	0.00
42 TM	1,2-Dichloroethane	0.556	0.562	-1.1	96	0.00
43 T	Isopropyl Acetate	1.141	1.133	0.7	94	0.00
44 TM	Trichloroethene	0.366	0.382	-4.4	100	0.00
45 C	1,2-Dichloropropane	0.433	0.444	-2.5#	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX060319\
 Data File : VX010025.D
 Acq On : 03 Jun 2019 10:28
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 03:59:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.255	0.255	0.0	98	0.00
47 T	Bromodichloromethane	0.485	0.529	-9.1	101	0.00
48 T	Methyl methacrylate	0.565	0.576	-1.9	94	0.00
49 T	1,4-Dioxane	0.011	0.010	9.1	92	0.00
50 S	Toluene-d8	1.277	1.284	-0.5	94	0.00
51 T	4-Methyl-2-Pentanone	0.742	0.749	-0.9	93	0.00
52 CM	Toluene	0.914	0.972	-6.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.582	0.626	-7.6	98	0.00
54 T	cis-1,3-Dichloropropene	0.634	0.680	-7.3	98	0.00
55 T	1,1,2-Trichloroethane	0.368	0.383	-4.1	98	0.00
56 T	Ethyl methacrylate	0.655	0.717	-9.5	97	0.00
57 T	1,3-Dichloropropane	0.660	0.675	-2.3	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.343	0.352	-2.6	92	0.00
59 T	2-Hexanone	0.571	0.597	-4.6	94	0.00
60 T	Dibromochloromethane	0.362	0.404	-11.6	101	0.00
61 T	1,2-Dibromoethane	0.378	0.396	-4.8	99	0.00
62 S	4-Bromofluorobenzene	0.459	0.480	-4.6	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.349	0.381	-9.2	102	0.00
65 PM	Chlorobenzene	1.047	1.095	-4.6	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.395	-8.8	100	0.00
67 C	Ethyl Benzene	1.879	2.048	-9.0#	101	0.00
68 T	m/p-Xylenes	0.694	0.759	-9.4	101	0.00
69 T	o-Xylene	0.666	0.733	-10.1	101	0.00
70 T	Styrene	1.168	1.284	-9.9	100	0.00
71 P	Bromoform	0.297	0.347	-16.8	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.689	3.865	-4.8	101	0.00
74 T	N-amyl acetate	2.204	2.226	-1.0	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.376	1.351	1.8	98	0.00
76 T	1,2,3-Trichloropropane	1.196	1.195	0.1	98	0.00
77 T	Bromobenzene	0.940	0.973	-3.5	105	0.00
78 T	n-propylbenzene	4.252	4.579	-7.7	102	0.00
79 T	2-Chlorotoluene	2.598	2.648	-1.9	100	0.00
80 T	1,3,5-Trimethylbenzene	3.108	3.304	-6.3	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.483	0.502	-3.9	99	0.00
82 T	4-Chlorotoluene	3.024	3.121	-3.2	101	0.00
83 T	tert-Butylbenzene	2.973	3.199	-7.6	103	0.00
84 T	1,2,4-Trimethylbenzene	3.156	3.330	-5.5	101	0.00
85 T	sec-Butylbenzene	3.537	3.866	-9.3	103	0.00
86 T	p-Isopropyltoluene	3.182	3.564	-12.0	104	0.00
87 T	1,3-Dichlorobenzene	1.674	1.770	-5.7	106	0.00
88 T	1,4-Dichlorobenzene	1.707	1.761	-3.2	105	0.00
89 T	n-Butylbenzene	2.990	3.361	-12.4	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX060319\
Data File : VX010025.D
Acq On : 03 Jun 2019 10:28
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 04 03:59:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.546	0.614	-12.5	106	0.00
91 T	1,2-Dichlorobenzene	1.650	1.721	-4.3	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.324	0.306	5.6	95	0.00
93 T	1,2,4-Trichlorobenzene	1.179	1.292	-9.6	108	0.00
94 T	Hexachlorobutadiene	0.592	0.662	-11.8	109	0.00
95 T	Naphthalene	3.670	3.857	-5.1	102	0.00
96 T	1,2,3-Trichlorobenzene	1.180	1.247	-5.7	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6