

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060320\
 Data File : VX016558.D
 Acq On : 03 Jun 2020 11:10
 Operator : JC/SP
 Sample : VX0603WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBL01

Quant Time: Jun 04 08:47:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	269895	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	432604	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	423100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	197512	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	155660	46.44	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.88%	
35) Dibromofluoromethane	5.46	113	126951	47.66	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.32%	
50) Toluene-d8	8.70	98	530463	51.16	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	11.12	95	208426	52.03	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.06%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060320\
Data File : VX016558.D
Acq On : 03 Jun 2020 11:10
Operator : JC/SP
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Quant Time: Jun 04 08:47:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

