

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060320\
 Data File : VX016565.D
 Acq On : 03 Jun 2020 14:49
 Operator : JC/SP
 Sample : L2867-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP20-CHY02

Quant Time: Jun 04 09:00:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	252462	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	410410	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	409087	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206499	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	150112	47.88	ug/l	0.00
Spiked Amount						
			Recovery	=	95.76%	
35) Dibromofluoromethane	5.47	113	121951	48.26	ug/l	0.00
Spiked Amount						
			Recovery	=	96.52%	
50) Toluene-d8	8.70	98	506180	51.46	ug/l	0.00
Spiked Amount						
			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.12	95	202977	53.41	ug/l	0.00
Spiked Amount						
			Recovery	=	106.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	20088	14.624	ug/l	92
17) Carbon Disulfide	2.56	76	8168	1.132	ug/l #	94
25) 2-Butanone	4.64	43	5584	2.399	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	5458	1.756	ug/l	83
37) Ethyl Acetate	4.79	43	13854	2.970	ug/l	98
44) Trichloroethene	7.19	130	13684	3.727	ug/l	83
59) 2-Hexanone	9.48	43	1644	0.467	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060320\
Data File : VX016565.D
Acq On : 03 Jun 2020 14:49
Operator : JC/SP
Sample : L2867-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP20-CHY02

Quant Time: Jun 04 09:00:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

