

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060320\
 Data File : VX016570.D
 Acq On : 03 Jun 2020 16:43
 Operator : JC/SP
 Sample : L2867-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP20-MMW3414

Quant Time: Jun 04 09:10:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	248106	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	400839	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	404614	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	213251	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	146185	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
35) Dibromofluoromethane	5.47	113	119036	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
50) Toluene-d8	8.70	98	495033	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
62) 4-Bromofluorobenzene	11.12	95	203630	54.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.72%	
Target Compounds						
4) Vinyl Chloride	1.39	62	1550	0.497	ug/l #	88
16) Acetone	2.42	43	5712	4.231	ug/l	100
27) cis-1,2-Dichloroethene	4.57	96	5445	1.782	ug/l	88
37) Ethyl Acetate	4.80	43	12171	2.671	ug/l	99
44) Trichloroethene	7.18	130	2285	0.637	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060320\
Data File : VX016570.D
Acq On : 03 Jun 2020 16:43
Operator : JC/SP
Sample : L2867-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP20-MMW3414

Quant Time: Jun 04 09:10:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

