

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022383.D
 Acq On : 03 Jun 2021 22:16
 Operator : JC/MD
 Sample : VX0603WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 04 03:56:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	76828	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	161437	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	145632	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64246	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	66960	56.50	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	=	113.00%	
35) Dibromofluoromethane	5.397	113	50295	49.13	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	98.26%	
50) Toluene-d8	8.653	98	192567	48.83	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	=	97.66%	
62) 4-Bromofluorobenzene	11.085	95	71637	49.15	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	=	98.30%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	19784	21.18	ug/l	99
3) Chloromethane	1.294	50	23084	22.91	ug/l	97
4) Vinyl Chloride	1.374	62	23404	21.95	ug/l	97
5) Bromomethane	1.599	94	17910	28.82	ug/l	96
6) Chloroethane	1.685	64	13682	21.55	ug/l	94
7) Trichlorofluoromethane	1.886	101	32550	21.24	ug/l	98
8) Diethyl Ether	2.142	74	13422	20.36	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.331	101	19347	21.26	ug/l	90
10) Methyl Iodide	2.453	142	22335	22.83	ug/l	94
11) Tert butyl alcohol	2.977	59	33979	125.36	ug/l	100
12) 1,1-Dichloroethene	2.319	96	19186	21.79	ug/l	96
13) Acrolein	2.239	56	20051	111.53	ug/l	100
14) Allyl chloride	2.666	41	35862	22.17	ug/l	97
15) Acrylonitrile	3.075	53	67131	125.80	ug/l	99
16) Acetone	2.386	43	62112	120.36	ug/l	89
17) Carbon Disulfide	2.514	76	44534	19.31	ug/l	96
18) Methyl Acetate	2.715	43	33572	24.46	ug/l	90
19) Methyl tert-butyl Ether	3.123	73	68090	22.57	ug/l	99
20) Methylene Chloride	2.794	84	24508	21.39	ug/l	# 88
21) trans-1,2-Dichloroethene	3.093	96	22044	23.02	ug/l	90
22) Diisopropyl ether	3.776	45	71730	23.03	ug/l	94
23) Vinyl Acetate	3.733	43	348900	116.33	ug/l	98
24) 1,1-Dichloroethane	3.617	63	41688	22.81	ug/l	99
25) 2-Butanone	4.574	43	104152	124.76	ug/l	92
26) 2,2-Dichloropropane	4.489	77	26086	16.08	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	26551	22.80	ug/l	97
28) Bromochloromethane	4.904	49	21274	23.99	ug/l	84
29) Tetrahydrofuran	5.019	42	60598	121.95	ug/l	91
30) Chloroform	5.105	83	41931	22.20	ug/l	92
31) Cyclohexane	5.477	56	33909	22.77	ug/l	94
32) 1,1,1-Trichloroethane	5.391	97	35210	21.82	ug/l	98
36) 1,1-Dichloropropene	5.702	75	31242	20.93	ug/l	97
37) Ethyl Acetate	4.727	43	39258	21.45	ug/l	94
38) Carbon Tetrachloride	5.684	117	28201	19.31	ug/l	94
39) Methylcyclohexane	7.385	83	33404	19.73	ug/l	99
40) Benzene	6.044	78	98589	21.46	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022383.D
 Acq On : 03 Jun 2021 22:16
 Operator : JC/MD
 Sample : VX0603WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 04 03:56:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	20572	21.01	ug/l	94
42) 1,2-Dichloroethane	6.098	62	34252	21.29	ug/l	99
43) Isopropyl Acetate	6.355	43	63972	21.37	ug/l	94
44) Trichloroethene	7.129	130	23021	20.17	ug/l	98
45) 1,2-Dichloropropane	7.440	63	25133	21.45	ug/l	96
46) Dibromomethane	7.586	93	16826	21.17	ug/l #	84
47) Bromodichloromethane	7.824	83	29382	18.47	ug/l	97
48) Methyl methacrylate	7.702	41	30112	21.83	ug/l	88
49) 1,4-Dioxane	7.665	88	12897	480.21	ug/l	97
51) 4-Methyl-2-Pentanone	8.580	43	205658	114.70	ug/l	91
52) Toluene	8.726	92	59791	20.56	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	32864	17.90	ug/l	96
54) cis-1,3-Dichloropropene	8.372	75	37566	19.03	ug/l	96
55) 1,1,2-Trichloroethane	9.159	97	24643	20.87	ug/l	98
56) Ethyl methacrylate	9.122	69	39945	20.93	ug/l #	87
57) 1,3-Dichloropropane	9.311	76	42755	21.21	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	98709	95.63	ug/l	96
59) 2-Hexanone	9.433	43	156917	116.36	ug/l	90
60) Dibromochloromethane	9.525	129	20087	17.49	ug/l	99
61) 1,2-Dibromoethane	9.616	107	24494	20.83	ug/l	100
64) Tetrachloroethene	9.281	164	18945	19.81	ug/l	92
65) Chlorobenzene	10.086	112	62366	20.57	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.171	131	20764	19.71	ug/l	99
67) Ethyl Benzene	10.195	91	113112	20.28	ug/l	97
68) m/p-Xylenes	10.305	106	85071	40.36	ug/l	99
69) o-Xylene	10.646	106	41290	20.14	ug/l	98
70) Styrene	10.659	104	69053	20.56	ug/l	99
71) Bromoform	10.805	173	11516	15.27	ug/l #	98
73) Isopropylbenzene	10.963	105	108314	19.92	ug/l	100
74) N-amyl acetate	10.848	43	56276	21.80	ug/l	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	38156	21.17	ug/l	98
76) 1,2,3-Trichloropropane	11.244	75	43385	27.37	ug/l	88
77) Bromobenzene	11.201	156	22455	19.48	ug/l	84
78) n-propylbenzene	11.305	91	128991	20.25	ug/l	97
79) 2-Chlorotoluene	11.366	91	76079	19.69	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	92607	20.55	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.024	75	9550	14.28	ug/l	88
82) 4-Chlorotoluene	11.457	91	88523	19.89	ug/l	97
83) tert-Butylbenzene	11.719	119	84376	19.52	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	90459	19.94	ug/l	99
85) sec-Butylbenzene	11.890	105	109349	19.91	ug/l	97
86) p-Isopropyltoluene	12.012	119	90137	19.43	ug/l	97
87) 1,3-Dichlorobenzene	11.975	146	43239	19.45	ug/l	96
88) 1,4-Dichlorobenzene	12.043	146	44295	19.28	ug/l	96
89) n-Butylbenzene	12.335	91	82295	19.18	ug/l	98
90) Hexachloroethane	12.542	117	11938	15.57	ug/l	78
91) 1,2-Dichlorobenzene	12.341	146	42763	20.03	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	7538	19.55	ug/l	78
93) 1,2,4-Trichlorobenzene	13.591	180	24726	19.19	ug/l	96
94) Hexachlorobutadiene	13.725	225	8717	17.33	ug/l	97
95) Naphthalene	13.780	128	105106	20.92	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	25349	19.52	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022383.D
Acq On : 03 Jun 2021 22:16
Operator : JC/MD
Sample : VX0603WBS02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Jun 04 03:56:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jun 04 03:56:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 76828
Ion Ratio Lower Upper
168 100
99 69.4 45.8 68.6#



#2
Dichlorodifluoromethane
Concen: 21.18 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 85 Resp: 19784
Ion Ratio Lower Upper
85 100
87 32.2 16.4 49.2









#3
Chloromethane
Concen: 22.91 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 23084
Ion Ratio Lower Upper
50 100
52 31.1 26.4 39.6



#4
Vinyl Chloride
Concen: 21.95 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 62 Resp: 23404
Ion Ratio Lower Upper
62 100
64 30.2 25.7 38.5









#5
Bromomethane
Concen: 28.82 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.013 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 17910
Ion Ratio Lower Upper
94 100
96 91.3 76.4 114.6



#6
Chloroethane
Concen: 21.55 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 64 Resp: 13682
Ion Ratio Lower Upper
64 100
66 35.2 25.4 38.0









#7
Trichlorofluoromethane
Concen: 21.24 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 32550
Ion Ratio Lower Upper
101 100
103 67.9 53.2 79.8



#8
Diethyl Ether
Concen: 20.36 ug/l
RT: 2.142 min Scan# 173
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 74 Resp: 13422
Ion Ratio Lower Upper
74 100
45 105.2 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.26 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101	Resp:	19347
Ion	Ratio	Lower Upper
101	100	
85	46.2	34.9 52.3
151	65.3	61.4 92.2



#10
Methyl Iodide
Concen: 22.83 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:142	Resp:	22335
Ion	Ratio	Lower Upper
142	100	
127	46.8	33.3 49.9
141	13.9	11.4 17.0









#11

Tert butyl alcohol
Concen: 125.36 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.012 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 59 Resp: 33979
Ion Ratio Lower Upper
59 100
57 9.9 7.8 11.8




#12

1,1-Dichloroethene
Concen: 21.79 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 96 Resp: 19186
Ion Ratio Lower Upper
96 100
61 162.7 124.6 187.0
98 63.4 50.6 76.0









#13
Acrolein
Concen: 111.53 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 20051
Ion Ratio Lower Upper
56 100
55 70.2 56.4 84.6



#14
Allyl chloride
Concen: 22.17 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 41 Resp: 35862
Ion Ratio Lower Upper
41 100
39 64.8 51.6 77.4
76 33.1 30.5 45.7









#15
Acrylonitrile
Concen: 125.80 ug/l
RT: 3.075 min Scan# 326
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 67131
Ion Ratio Lower Upper
53 100
52 80.6 65.3 97.9
51 34.7 28.2 42.2



#16
Acetone
Concen: 120.36 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16


Tgt Ion: 43 Resp: 62112
Ion Ratio Lower Upper
43 100
58 28.6 27.8 41.8








#17
Carbon Disulfide
Concen: 19.31 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 76 Resp: 44534
Ion Ratio Lower Upper
76 100
78 10.5 7.2 10.8




#18
Methyl Acetate
Concen: 24.46 ug/l
RT: 2.715 min Scan# 267
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16



Tgt Ion: 43 Resp: 33572
Ion Ratio Lower Upper
43 100
74 22.2 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 22.57 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 68090
Ion Ratio Lower Upper
73 100
57 21.8 17.1 25.7



#20
Methylene Chloride
Concen: 21.39 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 84 Resp: 24508
Ion Ratio Lower Upper
84 100
49 126.4 88.6 132.8
51 42.9 27.0 40.6#
86 66.2 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 23.02 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 96 Resp: 22044
Ion Ratio Lower Upper
96 100
61 149.9 107.7 161.5
98 67.1 50.4 75.6



#22
Diisopropyl ether
Concen: 23.03 ug/l
RT: 3.776 min Scan# 441
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 45 Resp: 71730
Ion Ratio Lower Upper
45 100
43 68.1 58.0 87.0
87 25.5 25.4 38.0
59 12.7 10.2 15.2









#23
 Vinyl Acetate
 Concen: 116.33 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. 0.006 min
 Lab File: VX022383.D
 Acq: 03 Jun 2021 22:16




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 348900
 Ion Ratio Lower Upper
 43 100
 86 11.7 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 22.81 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX022383.D
 Acq: 03 Jun 2021 22:16

Tgt Ion: 63 Resp: 41688
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.8 11.2
 100 4.8 2.5 7.5









#25
2-Butanone
Concen: 124.76 ug/l
RT: 4.574 min Scan# 572
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 104152
Ion Ratio Lower Upper
43 100
72 25.3 23.7 35.5



#26
2,2-Dichloropropane
Concen: 16.08 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.012 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 77 Resp: 26086
Ion Ratio Lower Upper
77 100
97 25.4 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 22.80 ug/l

RT: 4.495 min Scan# 559





Delta R.T. 0.000 min

Lab File: VX022383.D

Acq: 03 Jun 2021 22:16

Tgt Ion: 96	Resp: 26551
Ion Ratio	Lower Upper
96 100	
61 138.9	0.0 285.6
98 62.2	0.0 129.6

Instrument :
MSVOA_X
ClientSampled :



#28

Bromochloromethane

Concen: 23.99 ug/l




RT: 4.904 min Scan# 626


Delta R.T. -0.006 min

Lab File: VX022383.D

Acq: 03 Jun 2021 22:16

Tgt Ion: 49	Resp: 21274
Ion Ratio	Lower Upper
49 100	
129 1.5	0.0 4.0
130 70.6	68.2 102.2





#29

Tetrahydrofuran

Concen: 121.95 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.006 min

Lab File: VX022383.D

Acq: 03 Jun 2021 22:16





Tgt Ion: 42 Resp: 60598

Ion Ratio Lower Upper

42 100

72 45.7 41.7 62.5

71 42.6 38.4 57.6



#30

Chloroform

Concen: 22.20 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX022383.D




Acq: 03 Jun 2021 22:16


Tgt Ion: 83 Resp: 41931

Ion Ratio Lower Upper

83 100


85 58.7 52.0 78.0








#31
Cyclohexane
Concen: 22.77 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion:	56	Resp:	33909
Ion	Ratio	Lower	Upper
56	100		
69	26.1	25.3	37.9
84	93.1	71.4	107.0




#32
1,1,1-Trichloroethane
Concen: 21.82 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16



Tgt Ion:	97	Resp:	35210
Ion	Ratio	Lower	Upper
97	100		
99	64.8	51.5	77.3
61	46.0	34.2	51.4





#33

1,2-Dichloroethane-d4

Concen: 56.50 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX022383.D





Acq: 03 Jun 2021 22:16

Tgt Ion: 65 Resp: 66960

Ion Ratio Lower Upper

65 100

67 53.4 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX022383.D

Acq: 03 Jun 2021 22:16




Tgt Ion: 114 Resp: 161437


Ion Ratio Lower Upper

114 100

63 21.8 0.0 40.8

88 17.3 0.0 32.8









#35
Dibromofluoromethane
Concen: 49.13 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 50295
Ion Ratio Lower Upper
113 100
111 101.6 81.9 122.9
192 15.9 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 20.93 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 75 Resp: 31242
Ion Ratio Lower Upper
75 100
110 34.5 18.1 54.4
77 30.5 25.3 37.9





#37

Ethyl Acetate

Concen: 21.45 ug/l





RT: 4.727 min Scan# 597

Delta R.T. 0.006 min

Lab File: VX022383.D

Acq: 03 Jun 2021 22:16

Tgt Ion:	43	Resp:	39258
Ion Ratio	Lower	Upper	
43	100		
61	13.1	12.8	19.2
70	11.1	10.5	15.7



#38

Carbon Tetrachloride

Concen: 19.31 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.006 min

Lab File: VX022383.D

Acq: 03 Jun 2021 22:16

Tgt Ion:	117	Resp:	28201
Ion Ratio	Lower	Upper	
117	100		
119	103.2	76.2	114.2
121	30.6	24.8	37.2









#39
Methylcyclohexane
Concen: 19.73 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 33404
Ion Ratio Lower Upper
83 100
55 74.4 58.9 88.3
98 48.3 38.1 57.1



#40
Benzene
Concen: 21.46 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 78 Resp: 98589
Ion Ratio Lower Upper
78 100
77 20.8 18.9 28.3









#41
Methacrylonitrile
Concen: 21.01 ug/l
RT: 4.934 min Scan# 631
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 41 Resp: 20572




Ion	Ratio	Lower	Upper
41	100		
39	57.7	42.5	63.7
67	74.8	65.0	97.6
52	32.1	25.3	37.9




#42
1,2-Dichloroethane
Concen: 21.29 ug/l
RT: 6.098 min Scan# 822
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 62 Resp: 34252

Ion	Ratio	Lower	Upper
62	100		
98	9.8	0.0	20.2









#43
Isopropyl Acetate
Concen: 21.37 ug/l
RT: 6.355 min Scan# 864
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 63972
Ion Ratio Lower Upper
43 100
61 20.1 18.2 27.4
87 13.2 12.5 18.7



#44
Trichloroethene
Concen: 20.17 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:130 Resp: 23021
Ion Ratio Lower Upper
130 100
95 100.8 0.0 198.0





#45

1,2-Dichloropropane

Concen: 21.45 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. 0.006 min

Lab File: VX022383.D





Acq: 03 Jun 2021 22:16

Tgt Ion: 63 Resp: 25133

Ion Ratio Lower Upper

63 100

65 30.0 25.7 38.5



#46

Dibromomethane

Concen: 21.17 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX022383.D

Acq: 03 Jun 2021 22:16




Tgt Ion: 93 Resp: 16826


Ion Ratio Lower Upper

93 100

95 81.5 66.8 100.2

174 75.9 83.1 124.7#









#47
Bromodichloromethane
Concen: 18.47 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 29382
Ion Ratio Lower Upper
83 100
85 61.8 51.4 77.0
127 8.2 6.8 10.2



#48
Methyl methacrylate
Concen: 21.83 ug/l
RT: 7.702 min Scan# 1085
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 41 Resp: 30112
Ion Ratio Lower Upper
41 100
69 83.6 76.0 114.0
39 64.5 44.6 67.0









#49
1,4-Dioxane
Concen: 480.21 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. -0.019 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 88 Resp: 12897
Ion Ratio Lower Upper
88 100
43 33.2 23.2 34.8
58 67.8 55.0 82.4



#50
Toluene-d8
Concen: 48.83 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 98 Resp: 192567
Ion Ratio Lower Upper
98 100
100 66.0 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 114.70 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 205658
Ion Ratio Lower Upper
43 100
58 38.5 35.4 53.0



#52
Toluene
Concen: 20.56 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 92 Resp: 59791
Ion Ratio Lower Upper
92 100
91 171.8 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 17.90 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 32864
Ion Ratio Lower Upper
75 100
77 33.5 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 19.03 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16


Tgt Ion: 75 Resp: 37566
Ion Ratio Lower Upper
75 100
77 30.4 25.7 38.5
39 47.0 35.1 52.7








#55
1,1,2-Trichloroethane
Concen: 20.87 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	97	Resp:	24643
Ion	Ratio	Lower	Upper
97	100		
83	87.1	69.3	103.9
85	59.9	44.3	66.5
99	64.2	50.3	75.5



#56
Ethyl methacrylate
Concen: 20.93 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:	69	Resp:	39945
Ion	Ratio	Lower	Upper
69	100		
41	66.7	46.8	70.2
39	40.4	24.7	37.1









#57

1,3-Dichloropropane
Concen: 21.21 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 76 Resp: 42755
Ion Ratio Lower Upper
76 100
78 31.9 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 95.63 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 63 Resp: 98709
Ion Ratio Lower Upper
63 100
106 27.2 23.5 35.3









#59
2-Hexanone
Concen: 116.36 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 156917
Ion Ratio Lower Upper
43 100
58 54.5 30.9 92.8



#60
Dibromochloromethane
Concen: 17.49 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 129 Resp: 20087
Ion Ratio Lower Upper
129 100
127 77.5 38.4 115.2









#61
1,2-Dibromoethane
Concen: 20.83 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 24494
Ion Ratio Lower Upper
107 100
109 95.1 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 49.15 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 95 Resp: 71637
Ion Ratio Lower Upper
95 100
174 64.5 0.0 154.6
176 60.4 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.061 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 145632
Ion Ratio Lower Upper
117 100
82 55.8 47.4 71.2
119 32.0 26.6 39.8



#64
Tetrachloroethene
Concen: 19.81 ug/l
RT: 9.281 min Scan# 1344
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:164 Resp: 18945
Ion Ratio Lower Upper
164 100
166 127.7 103.4 155.2
129 103.1 70.8 106.2
131 100.9 72.5 108.7









#65
Chlorobenzene
Concen: 20.57 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 62366
Ion Ratio Lower Upper
112 100
114 30.5 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 19.71 ug/l
RT: 10.171 min Scan# 1490
Delta R.T. 0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:131 Resp: 20764
Ion Ratio Lower Upper
131 100
133 93.0 47.7 143.1
119 64.4 32.1 96.5





#67

Ethyl Benzene

Concen: 20.28 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX022383.D





Acq: 03 Jun 2021 22:16

Tgt Ion: 91 Resp: 113112

Ion Ratio Lower Upper

91 100

106 32.6 24.6 37.0



#68

m/p-Xylenes

Concen: 40.36 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX022383.D




Acq: 03 Jun 2021 22:16


Tgt Ion: 106 Resp: 85071

Ion Ratio Lower Upper

106 100

91 205.8 162.8 244.2









#69
o-Xylene
Concen: 20.14 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 41290
Ion Ratio Lower Upper
106 100
91 218.8 108.1 324.2



#70
Styrene
Concen: 20.56 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:104 Resp: 69053
Ion Ratio Lower Upper
104 100
78 50.8 40.0 60.0
103 53.9 43.4 65.0









#71
Bromoform
Concen: 15.27 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:173 Resp: 11516
Ion Ratio Lower Upper
173 100
175 49.3 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:152 Resp: 64246
Ion Ratio Lower Upper
152 100
115 77.5 41.1 123.3
150 169.2 0.0 349.0









#73
Isopropylbenzene
Concen: 19.92 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:105 Resp: 108314
Ion Ratio Lower Upper
105 100
120 26.2 13.2 39.6



#74
N-ethyl acetate
Concen: 21.80 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 43 Resp: 56276
Ion Ratio Lower Upper
43 100
70 41.2 39.0 58.6
55 23.5 21.5 32.3
61 23.5 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 21.17 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 38156
Ion Ratio Lower Upper
83 100
131 8.7 5.1 15.2
85 65.3 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 27.37 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 75 Resp: 43385
Ion Ratio Lower Upper
75 100
77 30.8 19.1 57.5









#77
Bromobenzene
Concen: 19.48 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:156 Resp: 22455
Ion Ratio Lower Upper
156 100
77 198.3 82.8 248.6
158 101.2 49.1 147.4



#78
n-propylbenzene
Concen: 20.25 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 91 Resp: 128991
Ion Ratio Lower Upper
91 100
120 21.7 11.5 34.5









#79
2-Chlorotoluene
Concen: 19.69 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 76079
Ion Ratio Lower Upper
91 100
126 31.7 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 20.55 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 105 Resp: 92607
Ion Ratio Lower Upper
105 100
120 46.6 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 14.28 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 9550
Ion Ratio Lower Upper
75 100
53 108.7 74.5 111.7
89 49.0 37.9 56.9



#82
4-Chlorotoluene
Concen: 19.89 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 91 Resp: 88523
Ion Ratio Lower Upper
91 100
126 28.1 14.9 44.5









#83
tert-Butylbenzene
Concen: 19.52 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:119 Resp: 84376
Ion Ratio Lower Upper
119 100
91 64.1 27.9 83.7
134 23.1 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 19.94 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:105 Resp: 90459
Ion Ratio Lower Upper
105 100
120 45.8 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 19.91 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.006 min
 Lab File: VX022383.D
 Acq: 03 Jun 2021 22:16




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 109349
 Ion Ratio Lower Upper
 105 100
 134 18.6 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 19.43 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022383.D
 Acq: 03 Jun 2021 22:16

Tgt Ion:119 Resp: 90137
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.7
 91 26.0 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 19.45 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 43239
Ion Ratio Lower Upper
146 100
111 45.3 20.2 60.5
148 62.3 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 19.28 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:146 Resp: 44295
Ion Ratio Lower Upper
146 100
111 44.7 19.6 58.8
148 63.0 31.6 95.0









#89
n-Butylbenzene
Concen: 19.18 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	91	Resp:	82295
Ion	Ratio	Lower	Upper
91	100		
92	54.5	27.4	82.0
134	22.9	13.2	39.5



#90
Hexachloroethane
Concen: 15.57 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.001 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:	117	Resp:	11938
Ion	Ratio	Lower	Upper
117	100		
201	60.9	40.4	121.2









#91
1,2-Dichlorobenzene
Concen: 20.03 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 42763
Ion Ratio Lower Upper
146 100
111 46.5 20.8 62.2
148 62.1 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.55 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion: 75 Resp: 7538
Ion Ratio Lower Upper
75 100
155 71.5 45.3 135.9
157 89.3 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 19.19 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:180 Resp: 24726
Ion Ratio Lower Upper
180 100
182 93.2 47.7 143.1
145 35.7 15.0 45.0



#94
Hexachlorobutadiene
Concen: 17.33 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.006 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:225 Resp: 8717
Ion Ratio Lower Upper
225 100
223 61.4 31.1 93.2
227 60.1 31.6 95.0









#95
Naphthalene
Concen: 20.92 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:128 Resp: 105106
Ion Ratio Lower Upper
128 100
127 12.2 10.3 15.5
129 10.8 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.52 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022383.D
Acq: 03 Jun 2021 22:16

Tgt Ion:180 Resp: 25349
Ion Ratio Lower Upper
180 100
182 95.7 47.6 142.9
145 35.5 16.1 48.3

