

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022384.D
 Acq On : 03 Jun 2021 22:40
 Operator : JC/MD
 Sample : VX0603WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 04 03:56:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	80528	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	158977	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64963	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	68962	55.52	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.04%
35) Dibromofluoromethane	5.391	113	52399	51.98	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.96%
50) Toluene-d8	8.653	98	201478	51.88	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.76%
62) 4-Bromofluorobenzene	11.085	95	76250	53.12	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.24%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.173	85	21398	21.86	ug/l 97
3) Chloromethane	1.294	50	23688	22.43	ug/l 98
4) Vinyl Chloride	1.374	62	24794	22.19	ug/l 98
5) Bromomethane	1.599	94	18789	28.84	ug/l 99
6) Chloroethane	1.685	64	15308	23.01	ug/l 99
7) Trichlorofluoromethane	1.886	101	35238	21.93	ug/l 97
8) Diethyl Ether	2.142	74	14267	20.65	ug/l 86
9) 1,1,2-Trichlorotrifluo...	2.331	101	20848	21.85	ug/l 90
10) Methyl Iodide	2.453	142	24351	23.75	ug/l 94
11) Tert butyl alcohol	2.977	59	36701	129.18	ug/l 99
12) 1,1-Dichloroethene	2.319	96	21613	23.42	ug/l 96
13) Acrolein	2.239	56	20111	106.73	ug/l 99
14) Allyl chloride	2.666	41	37954	22.39	ug/l 95
15) Acrylonitrile	3.075	53	69522	124.29	ug/l 99
16) Acetone	2.386	43	64170	118.64	ug/l 99
17) Carbon Disulfide	2.514	76	49107	20.31	ug/l 99
18) Methyl Acetate	2.709	43	34823	24.21	ug/l 92
19) Methyl tert-butyl Ether	3.123	73	71125	22.50	ug/l 99
20) Methylene Chloride	2.794	84	25250	21.03	ug/l 87
21) trans-1,2-Dichloroethene	3.093	96	24000	23.91	ug/l 92
22) Diisopropyl ether	3.770	45	76168	23.33	ug/l 89
23) Vinyl Acetate	3.733	43	370617	117.90	ug/l 95
24) 1,1-Dichloroethane	3.617	63	44891	23.43	ug/l 98
25) 2-Butanone	4.568	43	110555	126.34	ug/l 92
26) 2,2-Dichloropropane	4.483	77	28167	16.56	ug/l 98
27) cis-1,2-Dichloroethene	4.501	96	27634	22.63	ug/l 98
28) Bromochloromethane	4.904	49	22251	23.94	ug/l # 81
29) Tetrahydrofuran	5.026	42	64965	124.74	ug/l 90
30) Chloroform	5.105	83	45122	22.79	ug/l 100
31) Cyclohexane	5.477	56	36173	23.17	ug/l 93
32) 1,1,1-Trichloroethane	5.391	97	37540	22.19	ug/l 98
36) 1,1-Dichloropropene	5.702	75	33654	22.89	ug/l 98
37) Ethyl Acetate	4.721	43	43006	23.86	ug/l 94
38) Carbon Tetrachloride	5.684	117	29881	20.78	ug/l 96
39) Methylcyclohexane	7.391	83	35544	21.32	ug/l 98
40) Benzene	6.050	78	103620	22.91	ug/l 99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022384.D
 Acq On : 03 Jun 2021 22:40
 Operator : JC/MD
 Sample : VX0603WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 04 03:56:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	22311	23.14	ug/l	95
42) 1,2-Dichloroethane	6.098	62	36633	23.12	ug/l	98
43) Isopropyl Acetate	6.348	43	70033	23.76	ug/l	93
44) Trichloroethene	7.135	130	25167	22.39	ug/l	99
45) 1,2-Dichloropropane	7.440	63	26674	23.12	ug/l	97
46) Dibromomethane	7.586	93	17306	22.11	ug/l #	86
47) Bromodichloromethane	7.830	83	32130	20.51	ug/l	98
48) Methyl methacrylate	7.696	41	31265	23.02	ug/l	91
49) 1,4-Dioxane	7.665	88	14063	531.72	ug/l	97
51) 4-Methyl-2-Pentanone	8.580	43	218649	123.84	ug/l	91
52) Toluene	8.720	92	63315	22.10	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	36539	20.21	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	39461	20.30	ug/l #	90
55) 1,1,2-Trichloroethane	9.159	97	26076	22.42	ug/l	98
56) Ethyl methacrylate	9.122	69	41627	22.15	ug/l #	84
57) 1,3-Dichloropropane	9.311	76	45271	22.80	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.244	63	106495	104.77	ug/l	96
59) 2-Hexanone	9.433	43	166720	125.55	ug/l	89
60) Dibromochloromethane	9.525	129	21706	19.19	ug/l	95
61) 1,2-Dibromoethane	9.616	107	26430	22.83	ug/l	100
64) Tetrachloroethene	9.275	164	19858	20.68	ug/l	89
65) Chlorobenzene	10.086	112	65603	21.54	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	21797	20.60	ug/l	98
67) Ethyl Benzene	10.195	91	122749	21.91	ug/l	99
68) m/p-Xylenes	10.305	106	91391	43.17	ug/l	100
69) o-Xylene	10.646	106	44511	21.62	ug/l	98
70) Styrene	10.659	104	72755	21.57	ug/l	98
71) Bromoform	10.805	173	12438	16.32	ug/l #	98
73) Isopropylbenzene	10.963	105	117174	21.31	ug/l	100
74) N-amyl acetate	10.848	43	58558	22.44	ug/l	92
75) 1,1,2,2-Tetrachloroethane	11.213	83	40602	22.28	ug/l	98
76) 1,2,3-Trichloropropane	11.244	75	45554	28.42	ug/l	88
77) Bromobenzene	11.201	156	24242	20.79	ug/l	84
78) n-propylbenzene	11.305	91	137923	21.42	ug/l	98
79) 2-Chlorotoluene	11.366	91	80370	20.58	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	97294	21.35	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.024	75	11157	16.49	ug/l	91
82) 4-Chlorotoluene	11.457	91	95433	21.21	ug/l	97
83) tert-Butylbenzene	11.719	119	89809	20.55	ug/l	91
84) 1,2,4-Trimethylbenzene	11.756	105	97124	21.17	ug/l	99
85) sec-Butylbenzene	11.896	105	115675	20.83	ug/l	97
86) p-Isopropyltoluene	12.012	119	97605	20.81	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	45865	20.40	ug/l	97
88) 1,4-Dichlorobenzene	12.043	146	47466	20.43	ug/l	94
89) n-Butylbenzene	12.335	91	87984	20.28	ug/l	98
90) Hexachloroethane	12.542	117	13356	17.23	ug/l	72
91) 1,2-Dichlorobenzene	12.341	146	45650	21.14	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.945	75	7944	20.37	ug/l	81
93) 1,2,4-Trichlorobenzene	13.591	180	25991	19.95	ug/l	97
94) Hexachlorobutadiene	13.725	225	9621	18.92	ug/l	99
95) Naphthalene	13.780	128	112126	22.07	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	26298	20.03	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022384.D
Acq On : 03 Jun 2021 22:40
Operator : JC/MD
Sample : VX0603WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Jun 04 03:56:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jun 04 03:56:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 80528
Ion Ratio Lower Upper
168 100
99 71.0 45.8 68.6#



#2
Dichlorodifluoromethane
Concen: 21.86 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 85 Resp: 21398
Ion Ratio Lower Upper
85 100
87 31.1 16.4 49.2









#3
Chloromethane
Concen: 22.43 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 23688
Ion Ratio Lower Upper
50 100
52 34.1 26.4 39.6



#4
Vinyl Chloride
Concen: 22.19 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 62 Resp: 24794
Ion Ratio Lower Upper
62 100
64 31.0 25.7 38.5









#5
Bromomethane
Concen: 28.84 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.013 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 18789
Ion Ratio Lower Upper
94 100
96 94.4 76.4 114.6



#6
Chloroethane
Concen: 23.01 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 64 Resp: 15308
Ion Ratio Lower Upper
64 100
66 32.3 25.4 38.0









#7
 Trichlorofluoromethane
 Concen: 21.93 ug/l
 RT: 1.886 min Scan# 131
 Delta R.T. -0.000 min
 Lab File: VX022384.D
 Acq: 03 Jun 2021 22:40




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:101 Resp: 35238
 Ion Ratio Lower Upper
 101 100
 103 64.4 53.2 79.8



#8
 Diethyl Ether
 Concen: 20.65 ug/l
 RT: 2.142 min Scan# 173
 Delta R.T. 0.006 min
 Lab File: VX022384.D
 Acq: 03 Jun 2021 22:40

Tgt Ion: 74 Resp: 14267
 Ion Ratio Lower Upper
 74 100
 45 101.8 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.85 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:101 Resp: 20848
Ion Ratio Lower Upper
101 100
85 45.5 34.9 52.3
151 65.0 61.4 92.2



#10
Methyl Iodide
Concen: 23.75 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:142 Resp: 24351
Ion Ratio Lower Upper
142 100
127 46.8 33.3 49.9
141 14.0 11.4 17.0









#11

Tert butyl alcohol
Concen: 129.18 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.012 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 59 Resp: 36701
Ion Ratio Lower Upper
59 100
57 9.4 7.8 11.8




#12

1,1-Dichloroethene
Concen: 23.42 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40


Tgt Ion: 96 Resp: 21613
Ion Ratio Lower Upper
96 100
61 151.7 124.6 187.0
98 58.9 50.6 76.0








#13
Acrolein
Concen: 106.73 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 56 Resp: 20111
Ion Ratio Lower Upper
56 100
55 69.8 56.4 84.6



#14
Allyl chloride
Concen: 22.39 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40



Tgt Ion: 41 Resp: 37954
Ion Ratio Lower Upper
41 100
39 66.2 51.6 77.4
76 32.9 30.5 45.7









#15
Acrylonitrile
Concen: 124.29 ug/l
RT: 3.075 min Scan# 326
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 69522
Ion Ratio Lower Upper
53 100
52 82.1 65.3 97.9
51 35.5 28.2 42.2



#16
Acetone
Concen: 118.64 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40


Tgt Ion: 43 Resp: 64170
Ion Ratio Lower Upper
43 100
58 34.5 27.8 41.8








#17
Carbon Disulfide
Concen: 20.31 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion: 76 Resp: 49107
Ion Ratio Lower Upper
76 100
78 8.7 7.2 10.8



#18
Methyl Acetate
Concen: 24.21 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 43 Resp: 34823
Ion Ratio Lower Upper
43 100
74 23.4 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 22.50 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 71125
Ion Ratio Lower Upper
73 100
57 21.7 17.1 25.7



#20
Methylene Chloride
Concen: 21.03 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 84 Resp: 25250
Ion Ratio Lower Upper
84 100
49 130.7 88.6 132.8
51 39.8 27.0 40.6
86 64.9 50.3 75.5








#21
trans-1,2-Dichloroethene
Concen: 23.91 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 96 Resp: 24000
Ion Ratio Lower Upper
96 100
61 125.9 107.7 161.5
98 55.1 50.4 75.6



#22
Diisopropyl ether
Concen: 23.33 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 45 Resp: 76168
Ion Ratio Lower Upper
45 100
43 82.5 58.0 87.0
87 26.3 25.4 38.0
59 10.6 10.2 15.2









#23
 Vinyl Acetate
 Concen: 117.90 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. 0.006 min
 Lab File: VX022384.D
 Acq: 03 Jun 2021 22:40




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 370617
 Ion Ratio Lower Upper
 43 100
 86 10.4 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 23.43 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX022384.D
 Acq: 03 Jun 2021 22:40

Tgt Ion: 63 Resp: 44891
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.8 11.2
 100 4.9 2.5 7.5









#25
2-Butanone
Concen: 126.34 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 110555
Ion Ratio Lower Upper
43 100
72 25.1 23.7 35.5



#26
2,2-Dichloropropane
Concen: 16.56 ug/l
RT: 4.483 min Scan# 557
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 77 Resp: 28167
Ion Ratio Lower Upper
77 100
97 24.3 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 22.63 ug/l

RT: 4.501 min Scan# 560





Delta R.T. 0.006 min

Lab File: VX022384.D

Acq: 03 Jun 2021 22:40

Tgt Ion: 96 Resp: 27634

Ion	Ratio	Lower	Upper
96	100		
61	145.8	0.0	285.6
98	64.3	0.0	129.6



#28

Bromochloromethane

Concen: 23.94 ug/l

RT: 4.904 min Scan# 626




Delta R.T. -0.006 min


Lab File: VX022384.D

Acq: 03 Jun 2021 22:40

Tgt Ion: 49 Resp: 22251

Ion	Ratio	Lower	Upper
49	100		
129	1.6	0.0	4.0
130	67.6	68.2	102.2#





#29

Tetrahydrofuran

Concen: 124.74 ug/l

RT: 5.026 min Scan# 646





Delta R.T. 0.013 min

Lab File: VX022384.D

Acq: 03 Jun 2021 22:40

Tgt Ion:	42	Resp:	64965
Ion	Ratio	Lower	Upper
42	100		
72	45.1	41.7	62.5
71	40.5	38.4	57.6

Instrument :
MSVOA_X
ClientSampled :



#30

Chloroform

Concen: 22.79 ug/l




RT: 5.105 min Scan# 659


Delta R.T. -0.000 min

Lab File: VX022384.D

Acq: 03 Jun 2021 22:40

Tgt Ion:	83	Resp:	45122
Ion	Ratio	Lower	Upper
83	100		
85	65.0	52.0	78.0









#31
Cyclohexane
Concen: 23.17 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 36173
Ion Ratio Lower Upper
56 100
69 27.0 25.3 37.9
84 83.1 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 22.19 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 97 Resp: 37540
Ion Ratio Lower Upper
97 100
99 63.6 51.5 77.3
61 45.4 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 55.52 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX022384.D

Acq: 03 Jun 2021 22:40

Tgt Ion: 65 Resp: 68962

Ion Ratio Lower Upper





65 100

67 53.0 0.0 105.4

Instrument :

MSVOA_X

ClientSampled :



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX022384.D

Acq: 03 Jun 2021 22:40




Tgt Ion: 114 Resp: 158977

Ion Ratio Lower Upper

114 100

63 22.2 0.0 40.8

88 17.3 0.0 32.8









#35
Dibromofluoromethane
Concen: 51.98 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:113 Resp: 52399
Ion Ratio Lower Upper
113 100
111 104.3 81.9 122.9
192 15.8 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 22.89 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 75 Resp: 33654
Ion Ratio Lower Upper
75 100
110 34.6 18.1 54.4
77 30.6 25.3 37.9









#37
Ethyl Acetate
Concen: 23.86 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 43006
Ion Ratio Lower Upper
43 100
61 13.4 12.8 19.2
70 11.1 10.5 15.7



#38
Carbon Tetrachloride
Concen: 20.78 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 117 Resp: 29881
Ion Ratio Lower Upper
117 100
119 100.3 76.2 114.2
121 31.6 24.8 37.2









#39
Methylcyclohexane
Concen: 21.32 ug/l
RT: 7.391 min Scan# 1034
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 35544
Ion Ratio Lower Upper
83 100
55 75.3 58.9 88.3
98 49.5 38.1 57.1



#40
Benzene
Concen: 22.91 ug/l
RT: 6.050 min Scan# 814
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 78 Resp: 103620
Ion Ratio Lower Upper
78 100
77 23.1 18.9 28.3









#41
Methacrylonitrile
Concen: 23.14 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	41	Resp:	22311
Ion	Ratio	Lower	Upper
41	100		
39	53.2	42.5	63.7
67	72.3	65.0	97.6
52	31.2	25.3	37.9



#42
1,2-Dichloroethane
Concen: 23.12 ug/l
RT: 6.098 min Scan# 822
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:	62	Resp:	36633
Ion	Ratio	Lower	Upper
62	100		
98	9.4	0.0	20.2









#43
Isopropyl Acetate
Concen: 23.76 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 43 Resp: 70033
Ion Ratio Lower Upper
43 100
61 19.0 18.2 27.4
87 12.9 12.5 18.7



#44
Trichloroethene
Concen: 22.39 ug/l
RT: 7.135 min Scan# 992
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:130 Resp: 25167
Ion Ratio Lower Upper
130 100
95 98.4 0.0 198.0





#45

1,2-Dichloropropane

Concen: 23.12 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. 0.006 min

Lab File: VX022384.D





Acq: 03 Jun 2021 22:40

Tgt Ion: 63 Resp: 26674

Ion Ratio Lower Upper

63 100

65 30.6 25.7 38.5



#46

Dibromomethane

Concen: 22.11 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX022384.D

Acq: 03 Jun 2021 22:40




Tgt Ion: 93 Resp: 17306

Ion Ratio Lower Upper

93 100

95 84.0 66.8 100.2

174 78.7 83.1 124.7#









#47
Bromodichloromethane
Concen: 20.51 ug/l
RT: 7.830 min Scan# 1106
Delta R.T. 0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 32130
Ion Ratio Lower Upper
83 100
85 66.0 51.4 77.0
127 8.5 6.8 10.2



#48
Methyl methacrylate
Concen: 23.02 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 41 Resp: 31265
Ion Ratio Lower Upper
41 100
69 86.2 76.0 114.0
39 63.0 44.6 67.0





#49

1,4-Dioxane

Concen: 531.72 ug/l

RT: 7.665 min Scan# 1079





Delta R.T. -0.019 min

Lab File: VX022384.D

Acq: 03 Jun 2021 22:40

Tgt Ion: 88 Resp: 14063

Ion	Ratio	Lower	Upper
88	100		
43	31.0	23.2	34.8
58	66.7	55.0	82.4



#50

Toluene-d8

Concen: 51.88 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. -0.000 min

Lab File: VX022384.D


Acq: 03 Jun 2021 22:40

Tgt Ion: 98 Resp: 201478

Ion	Ratio	Lower	Upper
98	100		
100	65.7	51.9	77.9







Instrument :
MSVOA_X
ClientSampled :



#51
4-Methyl-2-Pentanone
Concen: 123.84 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 218649
Ion Ratio Lower Upper
43 100
58 38.1 35.4 53.0



#52
Toluene
Concen: 22.10 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 92 Resp: 63315
Ion Ratio Lower Upper
92 100
91 174.0 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 20.21 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 36539
Ion Ratio Lower Upper
75 100
77 31.9 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 20.30 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 75 Resp: 39461
Ion Ratio Lower Upper
75 100
77 29.7 25.7 38.5
39 52.9 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 22.42 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	97	Resp:	26076
Ion	Ratio	Lower	Upper
97	100		
83	85.8	69.3	103.9
85	58.9	44.3	66.5
99	62.5	50.3	75.5



#56
Ethyl methacrylate
Concen: 22.15 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:	69	Resp:	41627
Ion	Ratio	Lower	Upper
69	100		
41	68.9	46.8	70.2
39	41.2	24.7	37.1#









#57

1,3-Dichloropropane
Concen: 22.80 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 76 Resp: 45271
Ion Ratio Lower Upper
76 100
78 31.7 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 104.77 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 63 Resp: 106495
Ion Ratio Lower Upper
63 100
106 27.4 23.5 35.3








#59
2-Hexanone
Concen: 125.55 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 166720
Ion Ratio Lower Upper
43 100
58 53.8 30.9 92.8



#60
Dibromochloromethane
Concen: 19.19 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 129 Resp: 21706
Ion Ratio Lower Upper
129 100
127 81.0 38.4 115.2









#61
1,2-Dibromoethane
Concen: 22.83 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 26430
Ion Ratio Lower Upper
107 100
109 94.7 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 53.12 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 95 Resp: 76250
Ion Ratio Lower Upper
95 100
174 62.9 0.0 154.6
176 61.7 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 146288
Ion Ratio Lower Upper
117 100
82 60.5 47.4 71.2
119 33.0 26.6 39.8



#64
Tetrachloroethene
Concen: 20.68 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:164 Resp: 19858
Ion Ratio Lower Upper
164 100
166 119.3 103.4 155.2
129 102.9 70.8 106.2
131 99.1 72.5 108.7








#65
Chlorobenzene
Concen: 21.54 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 65603
Ion Ratio Lower Upper
112 100
114 33.4 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 20.60 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:131 Resp: 21797
Ion Ratio Lower Upper
131 100
133 96.1 47.7 143.1
119 66.7 32.1 96.5









#67
Ethyl Benzene
Concen: 21.91 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 122749
Ion Ratio Lower Upper
91 100
106 30.3 24.6 37.0



#68
m/p-Xylenes
Concen: 43.17 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 106 Resp: 91391
Ion Ratio Lower Upper
106 100
91 204.2 162.8 244.2









#69
o-Xylene
Concen: 21.62 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 44511
Ion Ratio Lower Upper
106 100
91 219.3 108.1 324.2



#70
Styrene
Concen: 21.57 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:104 Resp: 72755
Ion Ratio Lower Upper
104 100
78 52.8 40.0 60.0
103 54.7 43.4 65.0









#71
Bromoform
Concen: 16.32 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:173 Resp: 12438
Ion Ratio Lower Upper
173 100
175 46.9 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:152 Resp: 64963
Ion Ratio Lower Upper
152 100
115 77.0 41.1 123.3
150 163.2 0.0 349.0









#73
Isopropylbenzene
Concen: 21.31 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:105 Resp: 117174
Ion Ratio Lower Upper
105 100
120 26.4 13.2 39.6



#74
N-ethyl acetate
Concen: 22.44 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 43 Resp: 58558
Ion Ratio Lower Upper
43 100
70 42.1 39.0 58.6
55 24.1 21.5 32.3
61 24.7 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 22.28 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 40602
Ion Ratio Lower Upper
83 100
131 8.6 5.1 15.2
85 64.5 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 28.42 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 75 Resp: 45554
Ion Ratio Lower Upper
75 100
77 31.2 19.1 57.5









#77
Bromobenzene
Concen: 20.79 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:156 Resp: 24242
Ion Ratio Lower Upper
156 100
77 198.8 82.8 248.6
158 95.6 49.1 147.4



#78
n-propylbenzene
Concen: 21.42 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 91 Resp: 137923
Ion Ratio Lower Upper
91 100
120 21.8 11.5 34.5









#79
2-Chlorotoluene
Concen: 20.58 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 91 Resp: 80370
Ion Ratio Lower Upper
91 100
126 32.6 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 21.35 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 105 Resp: 97294
Ion Ratio Lower Upper
105 100
120 47.6 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 16.49 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 11157
Ion Ratio Lower Upper
75 100
53 102.9 74.5 111.7
89 42.5 37.9 56.9



#82
4-Chlorotoluene
Concen: 21.21 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 91 Resp: 95433
Ion Ratio Lower Upper
91 100
126 28.0 14.9 44.5









#83
tert-Butylbenzene
Concen: 20.55 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:119 Resp: 89809
Ion Ratio Lower Upper
119 100
91 63.8 27.9 83.7
134 24.1 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 21.17 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:105 Resp: 97124
Ion Ratio Lower Upper
105 100
120 44.0 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 20.83 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX022384.D
 Acq: 03 Jun 2021 22:40




Instrument :
 MSVOA_X
 ClientSampleId :

Tgt Ion:105 Resp: 115675
 Ion Ratio Lower Upper
 105 100
 134 18.6 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 20.81 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022384.D
 Acq: 03 Jun 2021 22:40

Tgt Ion:119 Resp: 97605
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.7
 91 25.5 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 20.40 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.007 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 45865
Ion Ratio Lower Upper
146 100
111 44.9 20.2 60.5
148 64.4 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 20.43 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:146 Resp: 47466
Ion Ratio Lower Upper
146 100
111 44.1 19.6 58.8
148 60.1 31.6 95.0









#89
n-Butylbenzene
Concen: 20.28 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 87984
Ion Ratio Lower Upper
91 100
92 54.1 27.4 82.0
134 24.1 13.2 39.5



#90
Hexachloroethane
Concen: 17.23 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.001 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 117 Resp: 13356
Ion Ratio Lower Upper
117 100
201 56.0 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 21.14 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:146 Resp: 45650
Ion Ratio Lower Upper
146 100
111 45.8 20.8 62.2
148 63.2 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.37 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion: 75 Resp: 7944
Ion Ratio Lower Upper
75 100
155 74.5 45.3 135.9
157 93.2 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 19.95 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:180 Resp: 25991
Ion Ratio Lower Upper
180 100
182 94.9 47.7 143.1
145 35.2 15.0 45.0



#94
Hexachlorobutadiene
Concen: 18.92 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.006 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:225 Resp: 9621
Ion Ratio Lower Upper
225 100
223 63.2 31.1 93.2
227 62.7 31.6 95.0









#95
Naphthalene
Concen: 22.07 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128 Resp: 112126
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.4 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 20.03 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022384.D
Acq: 03 Jun 2021 22:40

Tgt Ion:180 Resp: 26298
Ion Ratio Lower Upper
180 100
182 96.2 47.6 142.9
145 37.6 16.1 48.3

