

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022352.D
 Acq On : 03 Jun 2021 09:42
 Operator : JC/MD
 Sample : VX0603MBL01
 Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603MBL01

Quant Time: Jun 03 13:02:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	66460	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	142068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58368	56.934	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.860%	
35) Dibromofluoromethane	5.391	113	43947	48.780	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.560%	
50) Toluene-d8	8.653	98	174132	50.175	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.360%	
62) 4-Bromofluorobenzene	11.085	95	58219	45.386	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022352.D
Acq On : 03 Jun 2021 09:42
Operator : JC/MD
Sample : VX0603MBL01
Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603MBL01

Quant Time: Jun 03 13:02:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

