

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022358.D
 Acq On : 03 Jun 2021 12:02
 Operator : JC/MD
 Sample : M2560-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S1-D-TOP

Quant Time: Jun 03 13:05:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	62989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	137535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55775	57.402	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.800%
35) Dibromofluoromethane	5.397	113	40999	47.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.020%
50) Toluene-d8	8.653	98	169137	50.342	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.680%
62) 4-Bromofluorobenzene	11.085	95	60033	48.343	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.680%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	41260	97.522	ug/l	97
20) Methylene Chloride	2.794	84	8130	8.655	ug/l	96
43) Isopropyl Acetate	6.348	43	11416	4.477	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022358.D
Acq On : 03 Jun 2021 12:02
Operator : JC/MD
Sample : M2560-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S1-D-TOP

Quant Time: Jun 03 13:05:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

