

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022359.D
 Acq On : 03 Jun 2021 12:26
 Operator : JC/MD
 Sample : M2560-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S1-D-BOTTOM

Quant Time: Jun 03 13:05:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	62343	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	130732	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	119828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53240	55.361	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.720%
35) Dibromofluoromethane	5.397	113	39957	48.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.400%
50) Toluene-d8	8.653	98	163850	51.306	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.620%
62) 4-Bromofluorobenzene	11.085	95	57893	49.045	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.100%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	47815	114.187	ug/l	99
20) Methylene Chloride	2.794	84	7896	8.493	ug/l	# 77
43) Isopropyl Acetate	6.348	43	12481	5.149	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022359.D
Acq On : 03 Jun 2021 12:26
Operator : JC/MD
Sample : M2560-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S1-D-BOTTOM

Quant Time: Jun 03 13:05:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

