

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022361.D
 Acq On : 03 Jun 2021 13:15
 Operator : JC/MD
 Sample : M2572-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 91BR-20210528DL

Quant Time: Jun 04 03:49:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	59292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	124684	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112847	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49957	54.620	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.240%
35) Dibromofluoromethane	5.391	113	37827	47.841	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.680%
50) Toluene-d8	8.653	98	152454	50.054	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.100%
62) 4-Bromofluorobenzene	11.085	95	55984	49.729	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.460%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	523	0.726	ug/l	94
4) Vinyl Chloride	1.374	62	18764	22.807	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	519	0.702	ug/l #	65
27) cis-1,2-Dichloroethene	4.495	96	61451	68.362	ug/l	91
44) Trichloroethene	7.135	130	936	1.062	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022361.D
Acq On : 03 Jun 2021 13:15
Operator : JC/MD
Sample : M2572-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
91BR-20210528DL

Quant Time: Jun 04 03:49:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

