

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022366.D
 Acq On : 03 Jun 2021 15:13
 Operator : JC/MD
 Sample : M2567-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 04 03:51:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	62038	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	133571	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53949	56.374	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.740%	
35) Dibromofluoromethane	5.398	113	40101	47.342	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.680%	
50) Toluene-d8	8.653	98	163432	50.088	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.180%	
62) 4-Bromofluorobenzene	11.086	95	55853	46.311	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.620%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022366.D
Acq On : 03 Jun 2021 15:13
Operator : JC/MD
Sample : M2567-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 04 03:51:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

