

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022367.D
 Acq On : 03 Jun 2021 15:36
 Operator : JC/MD
 Sample : M2567-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 04 03:51:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	129207	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43510	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53394	56.870	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.740%
35) Dibromofluoromethane	5.397	113	39400	48.086	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	8.653	98	159239	50.451	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.900%
62) 4-Bromofluorobenzene	11.085	95	55117	47.245	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022367.D
Acq On : 03 Jun 2021 15:36
Operator : JC/MD
Sample : M2567-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 04 03:51:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

