

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022369.D
 Acq On : 03 Jun 2021 16:23
 Operator : JC/MD
 Sample : M2577-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-1

Quant Time: Jun 04 03:52:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	58339	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	125246	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	112346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51455	57.177	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.360%
35) Dibromofluoromethane	5.397	113	37425	47.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.240%
50) Toluene-d8	8.652	98	152693	49.907	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.820%
62) 4-Bromofluorobenzene	11.085	95	54975	48.613	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.220%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	18346	46.819	ug/l	98
25) 2-Butanone	4.574	43	3971	6.264	ug/l #	87
51) 4-Methyl-2-Pentanone	8.573	43	4720	3.393	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022369.D
Acq On : 03 Jun 2021 16:23
Operator : JC/MD
Sample : M2577-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-1

Quant Time: Jun 04 03:52:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

