

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022371.D
 Acq On : 03 Jun 2021 17:10
 Operator : JC/MD
 Sample : M2577-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-3

Quant Time: Jun 04 03:52:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	135363	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48158	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53532	54.500	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.000%	
35) Dibromofluoromethane	5.397	113	39666	46.209	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.420%	
50) Toluene-d8	8.653	98	166291	50.289	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.580%	
62) 4-Bromofluorobenzene	11.085	95	58520	47.880	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.760%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12460	29.133	ug/l	100
25) 2-Butanone	4.580	43	2544	3.677	ug/l #	86
51) 4-Methyl-2-Pentanone	8.586	43	2786	1.853	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022371.D
Acq On : 03 Jun 2021 17:10
Operator : JC/MD
Sample : M2577-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-3

Quant Time: Jun 04 03:52:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

