

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022374.D
 Acq On : 03 Jun 2021 18:21
 Operator : JC/MD
 Sample : M2583-20 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 35-38

Quant Time: Jun 04 03:53:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59499	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	125837	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	109968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51116	55.693	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.380%
35) Dibromofluoromethane	5.397	113	38053	47.686	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.380%
50) Toluene-d8	8.653	98	150926	49.098	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.085	95	52597	46.292	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	3853	9.641	ug/l	90
52) Toluene	8.726	92	3623	1.598	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022374.D
Acq On : 03 Jun 2021 18:21
Operator : JC/MD
Sample : M2583-20 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
35-38

Quant Time: Jun 04 03:53:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

