

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022376.D
 Acq On : 03 Jun 2021 19:08
 Operator : JC/MD
 Sample : M2583-22 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 54-57

Quant Time: Jun 04 03:54:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59730	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	126774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52489	56.968	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.940%
35) Dibromofluoromethane	5.397	113	38425	47.796	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.600%
50) Toluene-d8	8.653	98	154907	50.020	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.040%
62) 4-Bromofluorobenzene	11.085	95	52570	45.926	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.860%
Target Compounds						
16) Acetone	2.392	43	2516	6.271	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022376.D
Acq On : 03 Jun 2021 19:08
Operator : JC/MD
Sample : M2583-22 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
54-57

Quant Time: Jun 04 03:54:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

