

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022382.D
 Acq On : 03 Jun 2021 21:53
 Operator : JC/MD
 Sample : VX0603WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBL02

Quant Time: Jun 04 03:55:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60815	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	123566	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39766	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48315	51.502	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.000%	
35) Dibromofluoromethane	5.397	113	37423	47.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.520%	
50) Toluene-d8	8.653	98	150711	49.929	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.860%	
62) 4-Bromofluorobenzene	11.085	95	51878	46.498	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022382.D
Acq On : 03 Jun 2021 21:53
Operator : JC/MD
Sample : VX0603WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL02

Quant Time: Jun 04 03:55:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

