

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022389.D
 Acq On : 04 Jun 2021 00:37
 Operator : JC/MD
 Sample : PB36780ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB36780ZHE#04

Quant Time: Jun 04 05:39:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	135761	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56068	56.388	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.780%	
35) Dibromofluoromethane	5.391	113	41844	48.603	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.200%	
50) Toluene-d8	8.653	98	169659	51.157	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.320%	
62) 4-Bromofluorobenzene	11.085	95	56760	46.304	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.600%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.959	59	3516	15.460	ug/l #	73
16) Acetone	2.386	43	40988	94.670	ug/l	94
18) Methyl Acetate	2.715	43	1575	1.368	ug/l #	87
20) Methylene Chloride	2.794	84	9171	9.540	ug/l	91
25) 2-Butanone	4.574	43	5700	8.138	ug/l	90
43) Isopropyl Acetate	6.349	43	13488	5.358	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022389.D
Acq On : 04 Jun 2021 00:37
Operator : JC/MD
Sample : PB36780ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB36780ZHE#04

Quant Time: Jun 04 05:39:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

