

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022393.D
 Acq On : 04 Jun 2021 02:11
 Operator : JC/MD
 Sample : PB36780ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB36780ZHE#08

Quant Time: Jun 04 05:40:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61557	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	131995	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	115005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54215	57.095	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.180%
35) Dibromofluoromethane	5.397	113	39374	47.039	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.080%
50) Toluene-d8	8.653	98	162536	50.408	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.820%
62) 4-Bromofluorobenzene	11.085	95	53608	44.981	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.960%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	11917	28.822	ug/l	99
18) Methyl Acetate	2.715	43	1690	1.537	ug/l	94
20) Methylene Chloride	2.800	84	2744	2.989	ug/l #	85
43) Isopropyl Acetate	6.354	43	11042	4.512	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022393.D
Acq On : 04 Jun 2021 02:11
Operator : JC/MD
Sample : PB36780ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB36780ZHE#08

Quant Time: Jun 04 05:40:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

