

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022401.D
 Acq On : 04 Jun 2021 05:20
 Operator : JC/MD
 Sample : PB36780ZHE#16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB36780ZHE#16

Quant Time: Jun 04 07:09:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	60705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	126446	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51851	55.372	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 110.740%		
35) Dibromofluoromethane	5.397	113	38339	47.813	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.620%		
50) Toluene-d8	8.653	98	156591	50.695	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.400%		
62) 4-Bromofluorobenzene	11.085	95	53648	46.990	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 93.980%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.959	59	2326	10.860	ug/l #	79
16) Acetone	2.386	43	30933	75.864	ug/l	97
18) Methyl Acetate	2.721	43	1469	1.355	ug/l #	87
20) Methylene Chloride	2.800	84	6778	7.487	ug/l	95
25) 2-Butanone	4.568	43	4473	6.781	ug/l #	76
43) Isopropyl Acetate	6.354	43	11698	4.990	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022401.D
Acq On : 04 Jun 2021 05:20
Operator : JC/MD
Sample : PB36780ZHE#16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB36780ZHE#16

Quant Time: Jun 04 07:09:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

