

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022402.D
 Acq On : 04 Jun 2021 05:43
 Operator : JC/MD
 Sample : PB36780ZHE#17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB36780ZHE#17

Quant Time: Jun 04 07:09:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	130382	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	118681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54561	58.042	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.080%	
35) Dibromofluoromethane	5.391	113	38862	47.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.000%	
50) Toluene-d8	8.653	98	161960	50.851	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.700%	
62) 4-Bromofluorobenzene	11.086	95	57261	48.640	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.280%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	3381	15.725	ug/l #	79
16) Acetone	2.386	43	38931	95.113	ug/l	94
18) Methyl Acetate	2.715	43	1628	1.496	ug/l #	75
20) Methylene Chloride	2.794	84	7990	8.792	ug/l	93
25) 2-Butanone	4.581	43	5458	8.242	ug/l	90
43) Isopropyl Acetate	6.349	43	12795	5.293	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022402.D
Acq On : 04 Jun 2021 05:43
Operator : JC/MD
Sample : PB36780ZHE#17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB36780ZHE#17

Quant Time: Jun 04 07:09:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

