

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022403.D
 Acq On : 04 Jun 2021 06:07
 Operator : JC/MD
 Sample : M2580-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B1(0-2)

Quant Time: Jun 04 07:10:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60721	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	127784	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53115	56.707	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.420%
35) Dibromofluoromethane	5.397	113	38582	47.612	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.220%
50) Toluene-d8	8.653	98	159830	51.202	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.400%
62) 4-Bromofluorobenzene	11.085	95	56113	48.634	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.260%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13869	34.005	ug/l	89
18) Methyl Acetate	2.715	43	2110	1.945	ug/l #	85
20) Methylene Chloride	2.794	84	3579	3.952	ug/l #	65
43) Isopropyl Acetate	6.354	43	4906	2.071	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022403.D
Acq On : 04 Jun 2021 06:07
Operator : JC/MD
Sample : M2580-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B1(0-2)

Quant Time: Jun 04 07:10:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

