

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
 Data File : VX022404.D
 Acq On : 04 Jun 2021 06:30
 Operator : JC/MD
 Sample : M2580-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B1(2-4)

Quant Time: Jun 04 07:10:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58640	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	127526	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44867	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51273	56.683	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.360%
35) Dibromofluoromethane	5.403	113	37660	46.568	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.140%
50) Toluene-d8	8.653	98	158488	50.875	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.760%
62) 4-Bromofluorobenzene	11.085	95	53482	46.447	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.900%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12219	31.023	ug/l	96
18) Methyl Acetate	2.721	43	1124	1.073	ug/l #	80
20) Methylene Chloride	2.794	84	3153	3.605	ug/l #	79
43) Isopropyl Acetate	6.355	43	10404	4.400	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060321\
Data File : VX022404.D
Acq On : 04 Jun 2021 06:30
Operator : JC/MD
Sample : M2580-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B1(2-4)

Quant Time: Jun 04 07:10:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

