

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
 Data File : VX029182.D
 Acq On : 03 Jun 2022 10:59
 Operator : JC/MD
 Sample : VX0603WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBL01

Quant Time: Jun 06 02:33:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	327833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	608027	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	586178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	266718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	211473	57.205	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	114.400%
35) Dibromofluoromethane	5.385	113	200076	49.540	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.080%
50) Toluene-d8	8.647	98	716535	48.509	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.020%
62) 4-Bromofluorobenzene	11.079	95	282739	49.518	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
Data File : VX029182.D
Acq On : 03 Jun 2022 10:59
Operator : JC/MD
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Quant Time: Jun 06 02:33:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

