

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
 Data File : VX029185.D
 Acq On : 03 Jun 2022 13:19
 Operator : JC/MD
 Sample : N3164-02DL 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 07BR-20220531DL

Quant Time: Jun 06 02:34:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	330959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	601171	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	584581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	277431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	207328	55.554	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.100%
35) Dibromofluoromethane	5.385	113	200055	50.100	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.200%
50) Toluene-d8	8.653	98	709314	48.568	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.140%
62) 4-Bromofluorobenzene	11.079	95	284346	50.368	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.740%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	7437	2.246	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	141482	35.253	ug/l	86
44) Trichloroethene	7.129	130	5878	0.609	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
Data File : VX029185.D
Acq On : 03 Jun 2022 13:19
Operator : JC/MD
Sample : N3164-02DL 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
07BR-20220531DL

Quant Time: Jun 06 02:34:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

