

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
 Data File : VX029186.D
 Acq On : 03 Jun 2022 13:42
 Operator : JC/MD
 Sample : N3164-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 06BR-20220531

Quant Time: Jun 06 02:34:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	327389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	597173	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	576375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	268426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206466	55.926	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.860%
35) Dibromofluoromethane	5.391	113	198338	50.003	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.000%
50) Toluene-d8	8.653	98	703562	48.497	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.000%
62) 4-Bromofluorobenzene	11.079	95	277004	49.396	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.800%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5687	3.912	ug/l	96
23) Vinyl Acetate	3.727	43	3930	0.531	ug/l	94
29) Tetrahydrofuran	5.026	42	3322	2.209	ug/l #	76
37) Ethyl Acetate	4.715	43	762581	154.171	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
Data File : VX029186.D
Acq On : 03 Jun 2022 13:42
Operator : JC/MD
Sample : N3164-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
06BR-20220531

Quant Time: Jun 06 02:34:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

