

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
 Data File : VX029193.D
 Acq On : 03 Jun 2022 16:26
 Operator : JC/MD
 Sample : N3164-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 32S-20220531

Quant Time: Jun 06 02:35:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	313850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	579241	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	570814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	264319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	205782	58.145	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.300%
35) Dibromofluoromethane	5.391	113	194641	50.590	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.180%
50) Toluene-d8	8.653	98	688517	48.929	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.085	95	274770	50.514	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.020%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	3401	2.440	ug/l	# 87
37) Ethyl Acetate	4.715	43	134455	28.024	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
Data File : VX029193.D
Acq On : 03 Jun 2022 16:26
Operator : JC/MD
Sample : N3164-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
32S-20220531

Quant Time: Jun 06 02:35:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

