

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
 Data File : VX029197.D
 Acq On : 03 Jun 2022 17:59
 Operator : JC/MD
 Sample : N3164-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 35BR-20220531

Quant Time: Jun 06 02:36:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	311909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	580467	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	564992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	255748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206181	58.621	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 117.240%			
35) Dibromofluoromethane	5.391	113	191102	49.565	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.120%			
50) Toluene-d8	8.653	98	685386	48.604	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.200%			
62) 4-Bromofluorobenzene	11.085	95	271820	49.866	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.740%			
Target Compounds						
16) Acetone	2.380	43	7059	5.097	ug/l	95
23) Vinyl Acetate	3.733	43	3291	0.467	ug/l	94
29) Tetrahydrofuran	5.025	42	2551	1.781	ug/l #	75
37) Ethyl Acetate	4.715	43	552222	114.856	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
Data File : VX029197.D
Acq On : 03 Jun 2022 17:59
Operator : JC/MD
Sample : N3164-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
35BR-20220531

Quant Time: Jun 06 02:36:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

