

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
 Data File : VX029199.D
 Acq On : 03 Jun 2022 18:45
 Operator : JC/MD
 Sample : N3164-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 37S-20220531

Quant Time: Jun 06 02:36:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	310044	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	583619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	571929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	267120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	209609	59.954	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	119.900%
35) Dibromofluoromethane	5.391	113	194563	50.190	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.380%
50) Toluene-d8	8.653	98	694324	48.972	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.940%
62) 4-Bromofluorobenzene	11.085	95	278306	50.780	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.560%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	5088	3.696	ug/l	95
29) Tetrahydrofuran	5.031	42	2364	1.660	ug/l #	69
37) Ethyl Acetate	4.721	43	132739	27.459	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
Data File : VX029199.D
Acq On : 03 Jun 2022 18:45
Operator : JC/MD
Sample : N3164-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
37S-20220531

Quant Time: Jun 06 02:36:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

