

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
 Data File : VX029202.D
 Acq On : 03 Jun 2022 19:55
 Operator : JC/MD
 Sample : N3022-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 48BR-20220525

Quant Time: Jun 06 02:37:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	302851	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	569657	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	562103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	263678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	204346	59.836	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 119.680%	
35) Dibromofluoromethane	5.385	113	188519	49.823	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.640%	
50) Toluene-d8	8.653	98	681031	49.211	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.420%	
62) 4-Bromofluorobenzene	11.085	95	270987	50.657	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.320%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	2160	0.713	ug/l	# 76
9) 1,1,2-Trichlorotrifluo...	2.331	101	5239	1.876	ug/l	95
12) 1,1-Dichloroethene	2.319	96	11504	4.215	ug/l	96
21) trans-1,2-Dichloroethene	3.099	96	1338	0.434	ug/l	# 92
24) 1,1-Dichloroethane	3.617	63	4265	0.841	ug/l	# 90
27) cis-1,2-Dichloroethene	4.495	96	58207	15.850	ug/l	87
44) Trichloroethene	7.129	130	1967608	466.122	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060322\
Data File : VX029202.D
Acq On : 03 Jun 2022 19:55
Operator : JC/MD
Sample : N3022-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
48BR-20220525

Quant Time: Jun 06 02:37:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

