

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046461.D
 Acq On : 03 Jun 2025 10:38
 Operator : JC/MD
 Sample : VX0603MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603MBL01

Quant Time: Jun 04 01:22:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	73865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	145814	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64154	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	70467	51.171	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.340%
35) Dibromofluoromethane	5.373	113	51990	49.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.020%
50) Toluene-d8	8.647	98	183415	50.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.940%
62) 4-Bromofluorobenzene	11.079	95	74990	53.793	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046461.D
Acq On : 03 Jun 2025 10:38
Operator : JC/MD
Sample : VX0603MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603MBL01

Quant Time: Jun 04 01:22:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

