

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046468.D
 Acq On : 03 Jun 2025 13:23
 Operator : JC/MD
 Sample : Q2186-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-680-682

Quant Time: Jun 04 01:26:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138414	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66991	51.330	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.660%
35) Dibromofluoromethane	5.385	113	49830	49.994	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.980%
50) Toluene-d8	8.647	98	172065	49.877	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.079	95	68507	51.770	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.540%
Target Compounds						
16) Acetone	2.380	43	3841	7.340	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046468.D
Acq On : 03 Jun 2025 13:23
Operator : JC/MD
Sample : Q2186-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-680-682

Quant Time: Jun 04 01:26:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

