

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046472.D
 Acq On : 03 Jun 2025 14:56
 Operator : JC/MD
 Sample : Q2159-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP05-MHO-WC

Quant Time: Jun 04 04:09:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54309	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66931	53.169	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.340%	
35) Dibromofluoromethane	5.385	113	49337	50.717	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.440%	
50) Toluene-d8	8.647	98	170992	50.786	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.580%	
62) 4-Bromofluorobenzene	11.079	95	66394	51.408	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.820%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	46552	92.231	ug/l	99
18) Methyl Acetate	2.703	43	13731	11.723	ug/l	100
20) Methylene Chloride	2.788	84	3616	3.739	ug/l	89
43) Isopropyl Acetate	6.355	43	3730	1.514	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046472.D
Acq On : 03 Jun 2025 14:56
Operator : JC/MD
Sample : Q2159-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP05-MHO-WC

Quant Time: Jun 04 04:09:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

