

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046474.D
 Acq On : 03 Jun 2025 15:43
 Operator : JC/MD
 Sample : Q2160-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP05-MHH-WC

Quant Time: Jun 04 01:29:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	67084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	137015	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59439	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67671	54.108	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.220%			
35) Dibromofluoromethane	5.385	113	48719	49.378	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.760%			
50) Toluene-d8	8.646	98	171029	50.083	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.160%			
62) 4-Bromofluorobenzene	11.079	95	70157	53.558	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 107.120%			
Target Compounds					Qvalue	
16) Acetone	2.379	43	22658	45.184	ug/l	98
18) Methyl Acetate	2.708	43	14976	12.870	ug/l	100
20) Methylene Chloride	2.782	84	3276	3.409	ug/l #	86
37) Ethyl Acetate	4.732	43	8291m	5.062	ug/l	
43) Isopropyl Acetate	6.354	43	6156	2.464	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046474.D
Acq On : 03 Jun 2025 15:43
Operator : JC/MD
Sample : Q2160-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP05-MHH-WC

Quant Time: Jun 04 01:29:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/04/2025
Supervised By : Mahesh Dadoda 06/04/2025

