

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046477.D
 Acq On : 03 Jun 2025 16:53
 Operator : JC/MD
 Sample : Q2173-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-400-CF-402B-COMP-25

Quant Time: Jun 04 01:30:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136404	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67012	54.059	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.120%			
35) Dibromofluoromethane	5.385	113	48166	49.036	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.080%			
50) Toluene-d8	8.647	98	171320	50.393	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.780%			
62) 4-Bromofluorobenzene	11.079	95	67985	52.133	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.260%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	30047	60.454	ug/l	96
18) Methyl Acetate	2.703	43	15526	13.462	ug/l	99
20) Methylene Chloride	2.782	84	3863	4.056	ug/l	87
37) Ethyl Acetate	4.739	43	6961	4.269	ug/l #	88
43) Isopropyl Acetate	6.348	43	5039	2.026	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046477.D
Acq On : 03 Jun 2025 16:53
Operator : JC/MD
Sample : Q2173-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-400-CF-402B-COMP-25

Quant Time: Jun 04 01:30:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

