

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046480.D
 Acq On : 03 Jun 2025 18:04
 Operator : JC/MD
 Sample : Q2172-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP06-MHQ

Quant Time: Jun 04 01:31:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	68587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	132091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	58182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67090	52.468	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.940%	
35) Dibromofluoromethane	5.385	113	49083	50.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.400%	
50) Toluene-d8	8.647	98	173377	51.234	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.460%	
62) 4-Bromofluorobenzene	11.079	95	69659	53.664	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.320%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	23225	45.300	ug/l	96
18) Methyl Acetate	2.703	43	15455	12.991	ug/l	100
20) Methylene Chloride	2.782	84	3178	3.235	ug/l	97
37) Ethyl Acetate	4.739	43	7924	4.882	ug/l #	96
43) Isopropyl Acetate	6.355	43	6619	2.673	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046480.D
Acq On : 03 Jun 2025 18:04
Operator : JC/MD
Sample : Q2172-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP06-MHQ

Quant Time: Jun 04 01:31:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

