

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060419\
 Data File : VX010056.D
 Acq On : 04 Jun 2019 12:49
 Operator : JC/SP
 Sample : K3165-06DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 13911DL

Quant Time: Jun 05 07:43:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	163471	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	260602	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	241818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107447	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
35) Dibromofluoromethane	5.49	113	78673	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
50) Toluene-d8	8.71	98	338084	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	11.13	95	119094	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
Target Compounds						
16) Acetone	2.45	43	119209	89.350	ug/l #	56
25) 2-Butanone	4.68	43	27446	13.204	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060419\
Data File : VX010056.D
Acq On : 04 Jun 2019 12:49
Operator : JC/SP
Sample : K3165-06DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
13911DL

Quant Time: Jun 05 07:43:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

