

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060420\
 Data File : VX016583.D
 Acq On : 04 Jun 2020 15:59
 Operator : JC/SP
 Sample : L2894-07 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1-DUP-0566-200602

Quant Time: Jun 05 07:20:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	226043	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	366323	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	366810	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	203822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	135604	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
35) Dibromofluoromethane	5.46	113	112762	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
50) Toluene-d8	8.70	98	443789	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
62) 4-Bromofluorobenzene	11.12	95	192333	56.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.40%	
Target Compounds						
4) Vinyl Chloride	1.39	62	1720	0.605	ug/l	97
20) Methylene Chloride	2.82	84	3030	1.191	ug/l #	94
27) cis-1,2-Dichloroethene	4.56	96	3263	1.172	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060420\
Data File : VX016583.D
Acq On : 04 Jun 2020 15:59
Operator : JC/SP
Sample : L2894-07 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1-DUP-0566-200602

Quant Time: Jun 05 07:20:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

