

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060420\
 Data File : VX016588.D
 Acq On : 04 Jun 2020 17:53
 Operator : JC/SP
 Sample : L2894-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012RW163-200601

Quant Time: Jun 05 07:32:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	230225	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	371667	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	375889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	191248	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	136936	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
35) Dibromofluoromethane	5.47	113	113222	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
50) Toluene-d8	8.70	98	458677	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
62) 4-Bromofluorobenzene	11.12	95	189654	55.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.20%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.39	62	8534	2.947	ug/l	100
12) 1,1-Dichloroethene	2.35	96	2494	1.126	ug/l	98
16) Acetone	2.42	43	4608	3.679	ug/l #	88
21) trans-1,2-Dichloroethene	3.15	96	3120	1.273	ug/l	83
27) cis-1,2-Dichloroethene	4.57	96	1540429	543.417	ug/l	89
44) Trichloroethene	7.20	130	3984	1.198	ug/l	90
65) Chlorobenzene	10.12	112	6959	0.908	ug/l	93
91) 1,2-Dichlorobenzene	12.38	146	2857	0.490	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060420\
Data File : VX016588.D
Acq On : 04 Jun 2020 17:53
Operator : JC/SP
Sample : L2894-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012RW163-200601

Quant Time: Jun 05 07:32:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

