

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060420\
 Data File : VX016593.D
 Acq On : 04 Jun 2020 19:46
 Operator : JC/SP
 Sample : L2894-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW201-200601

Quant Time: Jun 05 07:47:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	233999	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	377763	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	377293	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	142231	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
35) Dibromofluoromethane	5.47	113	114152	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
50) Toluene-d8	8.70	98	464565	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	11.12	95	186262	53.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.48%	
Target Compounds						
16) Acetone	2.42	43	6640	5.215	ug/l	99
27) cis-1,2-Dichloroethene	4.56	96	1708	0.593	ug/l	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060420\
Data File : VX016593.D
Acq On : 04 Jun 2020 19:46
Operator : JC/SP
Sample : L2894-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW201-200601

Quant Time: Jun 05 07:47:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

