

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
 Data File : VX046498.D
 Acq On : 04 Jun 2025 14:15
 Operator : JC/MD
 Sample : Q2200-01DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-02B-66-060325DL

Quant Time: Jun 05 01:54:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	62989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125027	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60283	51.335	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.660%
35) Dibromofluoromethane	5.385	113	45387	50.412	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.820%
50) Toluene-d8	8.646	98	155704	49.967	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	63770	53.350	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.700%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.373	62	716	0.823	ug/l	98
12) 1,1-Dichloroethene	2.312	96	1603	2.146	ug/l #	91
27) cis-1,2-Dichloroethene	4.489	96	15418	17.051	ug/l	88
44) Trichloroethene	7.122	130	36397	42.679	ug/l	97
64) Tetrachloroethene	9.268	164	517	0.616	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060425\
Data File : VX046498.D
Acq On : 04 Jun 2025 14:15
Operator : JC/MD
Sample : Q2200-01DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-02B-66-060325DL

Quant Time: Jun 05 01:54:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

