

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060519\
 Data File : VX010072.D
 Acq On : 05 Jun 2019 13:21
 Operator : JC/SP
 Sample : K3190-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-EB01-20190604

Quant Time: Jun 06 06:50:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	177339	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	282601	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	253944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113506	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117637	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
35) Dibromofluoromethane	5.50	113	86296	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
50) Toluene-d8	8.71	98	356668	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.13	95	121702	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
Target Compounds						
16) Acetone	2.44	43	7801	5.390	ug/l	98
52) Toluene	8.79	92	8964	1.736	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060519\
Data File : VX010072.D
Acq On : 05 Jun 2019 13:21
Operator : JC/SP
Sample : K3190-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-EB01-20190604

Quant Time: Jun 06 06:50:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

