

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX060519\
 Data File : VX010080.D
 Acq On : 05 Jun 2019 16:28
 Operator : JC/SP
 Sample : K3190-12 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW307D-20190604

Quant Time: Jun 06 07:06:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	170513	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	270339	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248879	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112369	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	111645	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
35) Dibromofluoromethane	5.50	113	82760	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	8.71	98	346570	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
62) 4-Bromofluorobenzene	11.13	95	117495	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060519\
Data File : VX010080.D
Acq On : 05 Jun 2019 16:28
Operator : JC/SP
Sample : K3190-12 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW307D-20190604

Quant Time: Jun 06 07:06:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

