

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060519\
 Data File : VX010086.D
 Acq On : 05 Jun 2019 18:48
 Operator : JC/SP
 Sample : K3190-15 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW304S-20190604

Quant Time: Jun 06 07:16:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178970	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	285036	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	257507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114322	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116820	46.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.42%	
35) Dibromofluoromethane	5.50	113	87274	48.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.40%	
50) Toluene-d8	8.71	98	362037	49.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.50%	
62) 4-Bromofluorobenzene	11.13	95	122354	46.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060519\
Data File : VX010086.D
Acq On : 05 Jun 2019 18:48
Operator : JC/SP
Sample : K3190-15 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW304S-20190604

Quant Time: Jun 06 07:16:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

